

INTERPRETATION OF RESULTS

Automated Parameters	
Wavelength	340 nm
Cuvette	1 cm light path
Reaction Temperature	37°C
Measurement	Against distilled w
Reaction Type	Kinetic test
Reaction Direction	Decreasing
Sample Volume	100 µl
Reagent Volume	1000 µl
Delay/Lag/time	60 Secs
Interval time	30 Secs
No. of Readings	04
Blank Absorbance limit	> 0.8
Factor	1746
Low Normal at 37°C	0 U/l
High Normal at 37°C	40 U/l
Linearity at 37°C	400 U/l

MANUAL ASSAY PROCEDURE

Pipette into Test Tubes.

Working Reagent	1000 µl
Sample	100 µl



Mix well and after 1 min 37°C at 340 nm read absorbance values and start timer simultaneously. Measure absorbance Decrease every 30 sec. during 2 minutes & calculate $\Delta A / \text{min}$ ($\Delta A / 30 \text{ sec} \times 2$).

SAMPLE DILUTIONS:

1. This method is linear upto a concentration of 400 U/l.
2. If the concentration exceeds this value, the sample should be diluted 1:1 with 0.9% saline solution.
3. Repeat assay. Multiply the result by 2.

SGPT (ALT) IFCC Kinetic UV Test

SGPT (ALT) IFCC TEST KIT



LINEARITY

The method is linear to a concentration of 400 U/L. If the concentration exceeds this value, the sample should be diluted 1:1 with 0.9% saline solution and reassayed.
Multiply the result by 2.

Limit of detection: 5U/L.

Corporate Office:-

C/609, 6th Floor, Mittal Commercial, Unit 2,
Hasan Pada Road, Chimatpada Marol, Andheri
(E), Mumbai - 400059.

Factory Address:-

G/F- 12/13, Evershine-2, Survey NO.
307/3/1, Balitha, N.H No 48, Vapi,
Valsad, Gujarat, 396191.