

Urine Strips - 2 Para (Glucose, Protein)

TEST PARAMETERS

- **Glucose**
- **Protein**

STORAGE

1. Storage at room temperature between 15 – 30° C (59-89 F) and out of direct sunlight.
2. Do not use after expiry date.
3. Do not refrigerate or freeze.
4. Store all test strips in the original bottle.
Do not remove the desiccant from bottle.
5. Close the bottle cap tightly after each use.

Urine Strips - 2 Para (Glucose, Protein)

Urine Strips - 2 Para



PERFORMANCE CHARACTERISTICS

Studies comparing the 2 Para Urinalysis Strip and other commercially available strips resulted in greater than 99% agreement with 60 urine samples.

Corporate Office:-

**C/609, 6th Floor, Mittal Commercial, Unit 2,
Hasan Pada Road, Chimatpada Marol, Andheri
(E), Mumbai - 400059.**

Factory Address:-

**G/F- 12/13, Evershine-2, Survey NO.
307/3/1, Balitha, N.H No 48, Vapi,
Valsad, Gujarat, 396191.**